



Industrie Service

# CERTIFICATE

The **ATP-Test Station**  
Of the **TÜV SÜD Industrie Service GmbH**

confirms to the company

**Comet System, s.r.o**  
**1.maje 1220**  
**CZE-75661 Roznov pod Radhostem**

because of the positive results of the tested temperature recorder,

**G0841**

According to the test-report No. **1087 TS** from the 14<sup>th</sup> of December 2011, that the temperature recorder fulfils the requirements of the standard DIN EN 12830 for the determination:


- Accuracy class: **1**
- environmental conditions: **D**
- application range: **transportation and storage**


The manufacturer is allowed to mark the temperature recorder with the following test-report number:


**ATP-MUC 1087 TS**

This certificate  
is valid from **14.12.2011**

This certificate  
is valid to: **14.12.2016**

  
ATP-Prüfstelle  
Rudolf Glück  
Munich, the **14.12.2011**



  
Technical expert  
Thorsten Bennert